



**CORRIGENDUM**

**Corrigendum for the e tender “Supply/Installation/Commissioning of Blood Gas Analyzer on Reagent Rental Basis for GMC Barmulla”, issued vide this office tender reference no, 15 GMCB OF 2023, dated:- 16-06-2023.**

After receiving many representations regarding altering some technical specification for a better competition, the below mentioned changes are hereby accepted by Faculty Incharge Central Diagnostic Laboratory.

<b>Point No.</b>	<b>Tender specification</b>	<b>Request for amendment</b>	<b>Response to query (Recommendation)</b>
1.	Integration with automated sample insertion (optional).	May please be deleted	<b>Accepted</b>
2.	Compatible software: Providing access to any networked analyzer to view results, monitor performance and deploy standard configurations.	Compatible software: Should have a provision to transfer the data to lab computer	<b>Accepted</b> Compatible software: providing access to any networked analyzer to view results, monitor performance and deploy standard configurations. Compatible software Should have a provision to transfer the data to lab computer
3.	Independent printout for results on low noise graphic thermal paper. Demographic look-up and patient verification options including date and time.	Independent printout for results on low noise graphic/ thermal paper.	<b>Accepted</b>
4.	Certification: The equipment should be US-FDA certified. User base in India and official confirmation of good performance will be given definite consideration	The equipment should be US-FDA certified. Declaration of conformity to European CE as per IVD 98/79/EG with amending directions 2007/47 EC.	<b>Accepted</b> (The equipment should be US-FDA / European CE certified)
5.	Instrument must be quoted considering 3 level quality control twice per day or it should have quality management software to sense, record and rectify errors on real-time basis.	Instrument should have tri level quality control twice daily with leavy Jennings plot and software to sense, record control values and rectify the errors so that patients results quality can be maintained.	<b>Accepted</b>



Government of Jammu and Kashmir

**DEPARTMENT OF HEALTH AND MEDICAL EDUCATION**  
**OFFICE OF THE PRINCIPAL**  
**GOVERNMENT MEDICAL COLLEGE BARAMULLA**



Other terms and conditions shall remain same as already advertised.

Sd/-  
**Dr. (Prof.) Ruby Reshi**  
Principal  
Govt. Medical College  
Baramulla

No: GMCB/PO/2023-24/957/1648-53

Dated: -28-06-2023

**Copy to the:**

1. *Administrative Secretary, Health & Medical Education Department, J&K, Jammu.*
2. *Director, Finance, Health & Medical Education Department, J&K, Jammu.*
3. *Joint Director, Information Department, Kashmir, Srinagar with the request to publish the NIT in two leading dailies of the UT.*
4. *Chief Accounts Officer GMC Baramulla for information.*
5. *Medical Superintendent Associated Hospital GMC Baramulla for information.*